

AUG 23 2002

### Summary of Safety and Effectiveness

Company Name: Nicolet Biomedical  
5225 Verona Road  
Madison, WI 53711

Contact: Glen Hermanson, Manager of Standards and Compliance  
Phone: 608 441-2065  
Fax: 608 441-2007

Summary Date: June 17, 2002

Trade Name: BraiNet

Common Name: Electrode Cap

Classification Name: 21 CFR 882.1320; Product Code: GXY

Predicate Device:

510(k) Number: K780045  
Manufacture: Electro-Cap International  
Trade Name: Electro-Cap, Infa-Cap

#### **1.0 Description of Device**

The BraiNet is used by licensed medical professionals to support placement of electroencephalograph (EEG) electrodes on the scalp. The electrodes connect to medical equipment in support of stimulation and recording.

The BraiNet is provided to the user non-sterile. The BraiNet is a single patient use, disposable devices.

#### **2.0 Intended Use**

The intended use of the Nicolet BraiNet is the same as the predicate Electro-Cap. The Nicolet BraiNet is placed on the scalp to support electroencephalograph (EEG) electrode placement.

### 3.0 Technological

The BraiNet is made from elastic and Velcro commercial garment grade materials.

### 4.0 Conclusions

The intended use and technology of the Nicolet BraiNet is substantially equivalent to the predicate device. No new questions of safety or effectiveness are raised.



Food and Drug Administration  
9200 Corporate Boulevard  
Rockville MD 20850

**AUG 23 2002**

Nicolet Biomedical, Inc.  
c/o Gary Syring  
Quality & Regulatory Associates, LLC  
800 Levanger Lane  
Stoughton, Wisconsin 53589

Re: K021986

Trade/Device Name: BraiNet  
Regulation Number: 882.1320  
Regulation Name: Cutaneous electrode  
Regulatory Class: Class II  
Product Code: GXY  
Dated: June 17, 2002  
Received: June 18, 2002

Dear Mr. Syring:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.


Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Page 2 – Mr. Gary Syring

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801 and additionally 21 CFR Part 809.10 for in vitro diagnostic devices), please contact the Office of Compliance at (301) 594-4659. Additionally, for questions on the promotion and advertising of your device, please contact the Office of Compliance at (301) 594-4639. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <http://www.fda.gov/cdrh/dsma/dsmamain.html>

Sincerely yours,

for Celia M. Witten, Ph.D., M.D.

Director

Division of General, Restorative  
and Neurological Devices

Office of Device Evaluation

Center for Devices and Radiological Health

Enclosure

510(k) Number (if known): K021986

Device Name: Nicolet BraiNet


Indications For Use:

The Nicolet BraiNet is placed on the scalp to support electroencephalograph (EEG) electrode placement.

(PLEASE: DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

\_\_\_\_\_  
Concurrence of CDRH, Office of Device Evaluation (ODE)

(Optional Format 3-10-98)

  
\_\_\_\_\_  
(Division Sign-Off)

Division of General, Restorative  
and Neurological Devices

510(k) Number K021986